

MagnetOs Granules Synthetic Bone Void Filler - Patient Information Leaflet

MagnetOs Granules is provided in the following configurations:

Product code	Volume	Size
703-004-AU	0.5cc	1-2mm
703-008-AU	1cc	1-2mm
703-012-AU	2cc	1-2mm
703-016-AU	5cc	1-2mm
703-017-AU		2-4mm
703-021-AU	10cc	1-2mm
703-022-AU		2-4mm
703-024-AU	15cc (2x 7.5cc)	2-4mm
703-026-AU	20cc (2x 10cc)	2-4mm

Intended purpose

MagnetOs Granules is an implant that fills bone defects. Bone defects are also called bone voids. These are areas where bone is missing. This can be in areas like arms, legs, spine and pelvis. The treatment is for bone voids caused by surgery or injury. Surgeons pack MagnetOs Granules into these bony voids in your skeleton. Over time, the granules resorb and are replaced with your own bone.

MagnetOs Granules can be used to fill bone defects in children and adults.

Product description and performance

MagnetOs Granules is a synthetic, resorbable material. It consists of 65-75% Tri-Calcium Phosphate (TCP, $\text{Ca}_3(\text{PO}_4)_2$) and 25-35% Hydroxyapatite (HA, $\text{Ca}_5(\text{PO}_4)_3\text{OH}$). The granules are porous and similar to human bone. The surface has very small needle-shaped features. These characteristics help your bone grow towards and into the implant. As the bone heals, new bone gradually replaces the granules, helping the bone void fill completely.

MagnetOs Granules comes sterile and is for single patient use only. The doctor decides how much material to implant based on the size of the bone defect.

Expected device lifetime

After MagnetOs Granules is implanted in your body, it begins to resorb. It is fully replaced by new bone. This can take up to 2 years after surgery.

Undesirable adverse effects

Your doctor will give you information about possible undesirable adverse effects of bone implants. Some of the most common may include:

- Incomplete, or no bone growth into the defect;
- Delayed or absent bone healing;
- Fracture of the implant;
- Inflammatory or allergic reactions;
- Pain or inflammation if the implant contacts a nerve.

If you have any unusual or increased pain, swelling, or redness around your surgery site, please contact your doctor.

Warnings and precautions

Warnings:

All surgeries carry risks. There are undesirable reactions that can occur during or after surgery. Bone grafting surgery may also cause undesirable reactions. Some of the most common may include:

- Pain;
- Bruises (haematomas);
- Swelling (oedema);
- Fluid accumulation;
- Inflammation;
- Superficial wound infection;
- Deep wound infection with or without inflammatory disease of the bone (osteomyelitis);
- Loss of alignment of the bone (reduction);
- Movement of the bone implant (implant migration);
- Bulging (protrusion) or displacement (dislodgment) of the implant;
- General complications associated with anaesthesia or surgery.

Precautions:

- MagnetOs Granules is for use by surgeons with bone grafting experience. This includes the use of rigid fixation techniques with plates and screws.
- MagnetOs Granules may look similar to bone in X-ray imaging. It may hide certain diseases. Your doctor should consider this when evaluating X-rays.
- The MagnetOs products are safe during medical tests like Magnetic Resonance Imaging (MRI) or Computerised Tomography (CT) scans.
- As with any bone grafting surgery, the result may not be successful in every case. It may be necessary to reoperate to remove or replace the implant. This can be due to medical conditions or failure of MagnetOs Granules.
- Keep the implant card you got after surgery in a safe, dark, and dry place.

Extra information

Your doctor may check on you after your surgery as part of routine check-ups.

Reporting a problem with the medical device

In the event of a serious incident that is related to the device, contact Kuros Biosciences B.V. (see contact details below). Also contact the Australian Therapeutic Goods Administration (TGA). Serious incidents include any change in your condition requiring unplanned medical or surgical intervention. To report to the TGA, follow the instructions at www.tga.gov.au.

Legal Manufacturer

Kuros Biosciences B.V.

Prof. Bronkhorstlaan 10, building 48

3723 MB Bilthoven

The Netherlands

website: www.kurosbio.com

email: cs.international@kurosbio.com